

## Selection and ordering Data

Description	Quantity for 2 years	Quantity for 5 years		Order No.
<b>Analyzer unit</b>				
O-ring for analyzer chamber: 180, 90, 60, 20 mm	2	4		<b>C71121-Z100-A99</b>
Chopper				
• With motor, for 1 IR channel (7MB2335-...)	1	1		<b>C79451-A3468-B515</b>
• With motor, for 2 IR channels (7MB2337-..., 7MB2338-...)	1	1		<b>C79451-A3468-B516</b>
<b>Electronics</b>				
Motherboard, with firmware	-	1	B)	<b>C79451-A3494-D501</b>
Keypad	1	1		<b>C79451-A3492-B605</b>
LCD module	1	1		<b>C79451-A3494-B16</b>
Connector filter	-	1	F)	<b>W75041-E5602-K2</b>
Line switch (portable analyzer)	-	1		<b>W75050-T1201-U101</b>
Fusible element 220 ... 240 V	2	4		<b>W79054-L1010-T630</b>
Fusible element 100 ... 120 V	2	4		<b>W79054-L1011-T125</b>
<b>Other</b>				
Safety filter (zero gas), internal	2	2		<b>A5E00059149</b>
Safety filter (sample gas), internal	2	3		<b>C79127-Z400-A1</b>
Pressure switch	1	2		<b>C79302-Z1210-A2</b>
Flowmeter	1	2		<b>C79402-Z560-T1</b>
Set of gaskets for sample gas pump	2	5	D)	<b>C79402-Z666-E20</b>
Condensation trap (for portable unit, in sheet steel enclosure)	1	2		<b>C79451-A3008-B43</b>
Filter (for portable unit, in sheet steel enclosure)	1	2		<b>C79451-A3008-B60</b>
Oxygen sensor	1	1		<b>C79451-A3458-B55</b>
Sample gas pump 50 Hz	1	1		<b>C79451-A3494-B10</b>
Sample gas pump 60 Hz	1	1		<b>C79451-A3494-B11</b>
Solenoid valve	1	1		<b>C79451-A3494-B33</b>

B) Subject to export regulations AL: N, ECCN: 3A991X

D) Subject to export regulations AL: 9I999, ECCN: N

F) Subject to export regulations AL: N, ECCN: EAR99H